



Accreditation No. 179
Accredited for compliance with ISO/IEC 17025 - Calibration
The results of the tests, calibrations and/or measurements included
in this document are traceable to Australian/national standards.
Reference equipment has been calibrated by the National
Measurement Institute or NATA Accredited Laboratories.
This document shall not be reproduced, except in full.

Test Report on Earth Enhancing Compound LPI SRIMPLUS-20

Report Number: 179/21/1789N

Report Date: 19 March 2021

Calibrated by:

Lech Bartnik

Telephone:

08 8292 0172

Checked by:

Lech Bartnik

Authorised Signatory

Page: 1 of 8

W/O: 20210264e

Contact:

Street Address:

Postal Address:

Telephone:

Calibration and Testing Services

41-55 Barnes Avenue, Marleston SA 5033

GPO Box 77, Adelaide SA 5001

(08) 8292 0166

CLIENT DETAILS

Reference: LPI_SRIMPLUS_ERT2021-001
Name: Lightning Protection International Pty Ltd.
Address: 49 Patriach Drive, Huntingfield TAS 7055
Attention: Dr Franco D'Alessandro

ITEM DETAILS

Date Received: 16 February 2021
Manufacturer: Lightning Protection International
Product Name: LPI Srimplus-20

TEST DETAILS

Work Order: 20210264e
Date of Test: 17 February - 24 February 2021 and 1st and 13th of March
Test Site: Enerven Marlestone
Nature of Test: Resistance measurement of 3 samples.
Test Method: As defined in IEC 62561-7. Measurements taken at 4 frequencies: 50, 95, 115 and 128Hz sinusoidal signal and at 2 test currents: 100 and 300mA. Additional tests carried out at 1kHz, 5kHz, 10kHz and 20kHz, 100mA.
Test Arrangements: See photo in Appendix 1
Conditions of Test: 20°C ±1°C, Humidity 45-60%
Samples prepared as per customer's instructions.
Reference Equipment: Multifunction Calibrator Wavetek 9100 S/N 30633, calibraton due date 30 March 2021
Reference Multimeter Fluke 8846A S/N1397037, calibration due date 12 October 2021
Results of Tests: Refer to tables.
Uncertainty of Measurement: Resistance: ±0.1%rdg ≤10Ω
Estimated for a 95% probability.
Test Pass Criteria: Resistivity <0.1Ωm.

Uncertainty

The stated uncertainties have been estimated for 95% confidence limits. Unless stated otherwise in the report a coverage factor of k=2 has been used.

The uncertainties apply at the time of measurement and at the stated 'Conditions of Test'. They do not consider drift after the calibration date nor do they take into account the environment and the conditions in which the instrument may be used.

Note: Resistivity calculations are not NATA endorsed.

RESULTS

Day/Date	Time	Current	Frequency	Resistance [Ω]			Calculated Resistivity [Ω m]			Test Result	
				Sample 1	Sample 2	Sample 3	Sample 1	Sample 2	Sample 3		
1 17/02/2021	11:05	0.1	50	1.113	0.897	1.193	0.080	0.065	0.086	Pass	
		0.1	95	0.932	0.850	1.196	0.067	0.061	0.086	Pass	
		0.1	115	0.946	0.874	1.196	0.068	0.063	0.086	Pass	
		0.1	128	0.938	0.847	1.205	0.068	0.061	0.087	Pass	
		0.3	50	0.928	0.882	1.121	0.067	0.064	0.081	Pass	
		0.3	95	0.914	0.864	1.103	0.066	0.062	0.079	Pass	
	+1	115	0.3	115	0.908	0.860	1.100	0.065	0.062	0.079	Pass
			0.3	128	0.918	0.854	1.099	0.066	0.062	0.079	Pass
		50	0.1	50	0.911	0.878	1.046	0.066	0.063	0.075	Pass
			0.1	95	0.890	0.860	1.031	0.064	0.062	0.074	Pass
		115	0.1	115	0.896	0.847	1.025	0.065	0.061	0.074	Pass
			0.1	128	0.893	0.846	1.022	0.064	0.061	0.074	Pass
+4	50	0.3	50	0.903	0.865	1.037	0.065	0.062	0.075	Pass	
		0.3	95	0.893	0.854	1.025	0.064	0.061	0.074	Pass	
		0.3	115	0.890	0.849	1.021	0.064	0.061	0.074	Pass	
		0.3	128	0.889	0.847	1.019	0.064	0.061	0.073	Pass	
		0.1	50	0.925	0.884	1.160	0.067	0.064	0.084	Pass	
		0.1	95	0.920	0.875	1.136	0.066	0.063	0.082	Pass	
	+8	115	0.1	115	0.915	0.872	1.131	0.066	0.063	0.081	Pass
			0.1	128	0.912	0.871	1.128	0.066	0.063	0.081	Pass
		50	0.3	50	0.922	0.884	1.157	0.066	0.064	0.083	Pass
			0.3	95	0.912	0.875	1.138	0.066	0.063	0.082	Pass
		115	0.3	115	0.909	0.872	1.132	0.065	0.063	0.082	Pass
			0.3	128	0.907	0.870	1.128	0.065	0.063	0.081	Pass
+8	50	0.1	50	1.061	1.101	1.221	0.076	0.079	0.088	Pass	
		0.1	95	1.046	1.089	1.204	0.075	0.078	0.087	Pass	
	115	0.1	115	1.041	1.083	1.197	0.075	0.078	0.086	Pass	
		0.1	128	1.037	1.080	1.193	0.075	0.078	0.086	Pass	
	50	0.3	50	1.063	1.101	1.219	0.077	0.079	0.088	Pass	

Report Number: 179/21/1789N

Checked:



2 18/02/2021	9:15	0.3	95	1.050	1.085	1.207	0.076	0.078	0.087	Pass
		0.3	115	1.046	1.080	1.202	0.075	0.078	0.087	Pass
		0.3	128	1.047	1.077	1.201	0.075	0.078	0.086	Pass
		0.1	50	1.115	1.198	1.276	0.080	0.086	0.092	Pass
		0.1	95	1.099	1.187	1.263	0.079	0.085	0.091	Pass
		0.1	115	1.094	1.182	1.256	0.079	0.085	0.090	Pass
		0.1	128	1.091	1.178	1.251	0.079	0.085	0.090	Pass
	17:30	0.3	50	1.113	1.202	1.272	0.080	0.087	0.092	Pass
		0.3	95	1.092	1.190	1.262	0.079	0.086	0.091	Pass
		0.3	115	1.088	1.185	1.253	0.078	0.085	0.090	Pass
		0.3	128	1.085	1.181	1.248	0.078	0.085	0.090	Pass
		0.1	50	1.098	1.133	1.168	0.079	0.082	0.084	Pass
		0.1	95	1.079	1.115	1.145	0.078	0.080	0.082	Pass
		0.1	115	1.074	1.109	1.138	0.077	0.080	0.082	Pass
		0.1	128	1.070	1.106	1.134	0.077	0.080	0.082	Pass
3 19/02/2021	9:15	0.3	50	1.133	1.133	1.168	0.082	0.082	0.084	Pass
		0.3	95	1.078	1.116	1.146	0.078	0.080	0.082	Pass
		0.3	115	1.072	1.110	1.138	0.077	0.080	0.082	Pass
		0.3	128	1.069	1.106	1.134	0.077	0.080	0.082	Pass
		0.1	50	1.000	1.008	1.026	0.072	0.073	0.074	Pass
		0.1	95	0.981	0.995	1.020	0.071	0.072	0.073	Pass
		0.1	115	0.975	0.991	1.004	0.070	0.071	0.072	Pass
	17:35	0.1	128	0.971	0.989	1.001	0.070	0.071	0.072	Pass
		0.3	50	1.007	1.007	1.026	0.073	0.072	0.074	Pass
		0.3	95	0.988	0.995	1.010	0.071	0.072	0.073	Pass
		0.3	115	0.982	0.991	1.004	0.071	0.071	0.072	Pass
		0.3	128	0.978	0.988	1.001	0.070	0.071	0.072	Pass
		0.1	50	1.008	1.026	1.001	0.073	0.074	0.072	Pass
		0.1	95	0.996	0.965	0.986	0.072	0.069	0.071	Pass
		0.1	115	0.989	0.961	0.981	0.071	0.069	0.071	Pass
0.1	128	0.985	0.959	0.978	0.071	0.069	0.070	0.070	Pass	
	0.3	50	1.004	1.026	1.000	0.072	0.074	0.072	Pass	
	0.3	95	0.984	0.965	0.981	0.071	0.069	0.071	Pass	



4 20/02/2021	14:50	0.3	115	0.978	0.961	0.976	0.070	0.069	0.070	Pass
		0.3	128	0.974	0.958	0.974	0.070	0.069	0.070	Pass
		0.1	50	0.993	0.943	0.989	0.071	0.068	0.071	Pass
		0.1	95	0.981	0.935	0.976	0.071	0.067	0.070	Pass
		0.1	115	0.977	0.931	0.972	0.070	0.067	0.070	Pass
		0.1	128	0.974	0.929	0.970	0.070	0.067	0.070	Pass
		0.3	50	0.993	0.945	0.988	0.072	0.068	0.071	Pass
		0.3	95	0.981	0.935	0.976	0.071	0.067	0.070	Pass
		0.3	115	0.976	0.932	0.972	0.070	0.067	0.070	Pass
		0.3	128	0.974	0.930	0.970	0.070	0.067	0.070	Pass
5 21/02/021	15:40	0.1	50	1.004	0.950	1.003	0.072	0.068	0.072	Pass
		0.1	95	0.993	0.938	0.992	0.071	0.068	0.071	Pass
		0.1	115	0.989	0.935	0.988	0.071	0.067	0.071	Pass
		0.1	128	0.987	0.933	0.986	0.071	0.067	0.071	Pass
		0.3	50	1.006	0.947	1.003	0.072	0.068	0.072	Pass
		0.3	95	0.995	0.938	0.992	0.072	0.068	0.071	Pass
		0.3	115	0.991	0.933	0.988	0.071	0.067	0.071	Pass
		0.3	128	0.989	0.933	0.986	0.071	0.067	0.071	Pass
		0.1	50	0.980	0.942	0.989	0.071	0.068	0.071	Pass
		0.1	95	0.970	0.934	0.993	0.070	0.067	0.072	Pass
6 22/02/2021	9:35	0.1	115	0.965	0.932	0.990	0.069	0.067	0.071	Pass
		0.1	128	0.962	0.930	0.988	0.069	0.067	0.071	Pass
		0.3	50	0.980	0.977	1.003	0.071	0.070	0.072	Pass
		0.3	95	0.965	0.936	0.994	0.069	0.067	0.072	Pass
		0.3	115	0.960	0.933	0.990	0.069	0.067	0.071	Pass
		0.3	128	0.957	0.931	0.988	0.069	0.067	0.071	Pass
		0.1	50	0.976	0.925	1.006	0.070	0.067	0.072	Pass
		0.1	95	0.961	0.918	0.996	0.069	0.066	0.072	Pass
		0.1	115	0.956	0.916	0.993	0.069	0.066	0.071	Pass
		0.1	128	0.953	0.915	0.991	0.069	0.066	0.071	Pass
	19:15	0.3	50	0.976	0.928	1.005	0.070	0.067	0.072	Pass
		0.3	95	0.961	0.921	0.996	0.069	0.066	0.072	Pass
		0.3	115	0.956	0.919	0.992	0.069	0.066	0.071	Pass



7 23/02/2021	8:45	0.3	128	0.954	0.917	0.991	0.069	0.066	0.071	Pass		
		0.1	50	0.990	0.931	1.000	0.071	0.067	0.072	Pass		
		0.1	95	0.975	0.925	0.992	0.070	0.067	0.071	Pass		
		0.1	115	0.970	0.923	0.989	0.070	0.066	0.071	Pass		
		0.1	128	0.968	0.921	0.987	0.070	0.066	0.071	Pass		
		0.3	50	0.982	0.932	1.002	0.071	0.067	0.072	Pass		
		0.3	95	0.969	0.926	0.993	0.070	0.067	0.071	Pass		
		0.3	115	0.964	0.923	0.990	0.069	0.066	0.071	Pass		
		0.3	128	0.961	0.922	0.988	0.069	0.066	0.071	Pass		
		0.1	50	0.990	0.933	0.983	0.071	0.067	0.071	Pass		
		0.1	95	0.981	0.926	0.976	0.071	0.067	0.070	Pass		
		0.1	115	0.976	0.923	0.977	0.070	0.066	0.070	Pass		
		0.1	128	0.973	0.922	0.971	0.070	0.066	0.070	Pass		
		0.3	50	0.990	0.935	0.993	0.071	0.067	0.072	Pass		
		0.3	95	0.975	0.927	1.000	0.070	0.067	0.072	Pass		
		0.3	115	0.970	0.925	1.003	0.070	0.067	0.072	Pass		
		8 24/02/2021	11:45	0.3	128	0.968	0.923	1.003	0.070	0.066	0.072	Pass
0.1	50			0.996	0.924	0.994	0.072	0.067	0.072	Pass		
0.1	95			0.975	0.919	0.987	0.070	0.066	0.071	Pass		
0.1	115			0.972	0.916	0.984	0.070	0.066	0.071	Pass		
0.1	128			0.969	0.915	0.983	0.070	0.066	0.071	Pass		
0.3	50			0.989	0.928	0.994	0.071	0.067	0.072	Pass		
0.3	95			0.977	0.919	0.986	0.070	0.066	0.071	Pass		
0.3	115			0.973	0.917	0.984	0.070	0.066	0.071	Pass		
0.3	128			0.970	0.916	0.983	0.070	0.066	0.071	Pass		
0.1	1kHz			0.920	0.821	0.906	0.066	0.059	0.065	Pass		
0.1	5kHz			0.866	0.780	0.872	0.062	0.056	0.063	Pass		
0.1	10kHz			0.837	0.756	0.854	0.060	0.054	0.061	Pass		
0.1	20kHz			0.807	0.732	0.834	0.058	0.053	0.060	Pass		
13/03/2021	11:30			0.1	50	0.897	0.769	0.911	0.065	0.055	0.066	Pass
				0.1	95	0.884	0.768	0.910	0.064	0.055	0.066	Pass
				0.1	115	0.882	0.768	0.910	0.064	0.055	0.065	Pass
				0.1	128	0.882	0.768	0.910	0.063	0.055	0.065	Pass
		0.1	128	0.882	0.768	0.910	0.063	0.055	0.065	Pass		



0.3	50	0.922	0.769	0.911	0.066	0.055	0.066	Pass
0.3	95	0.883	0.768	0.909	0.064	0.055	0.065	Pass
0.3	115	0.883	0.768	0.909	0.064	0.055	0.065	Pass
0.3	128	0.883	0.768	0.909	0.064	0.055	0.065	Pass
0.1	1kHz	0.887	0.762	0.895	0.064	0.055	0.064	Pass
0.1	10kHz	0.868	0.747	0.883	0.063	0.054	0.064	Pass

Overall Test Result: PASS.



APPENDIX 1.

The calibrator Wavetek 9100 was used as a current source and the Fluke 8846A 6½ digit multimeter was used to measure the AC voltage.



APPENDIX 2.

The resistivity of the samples was calculated with the formula:

$$\rho = \frac{RA}{L}$$

Where: R- resistance masured, A – cross-sectional area of the sample between sensing electrodes, L – spacing between the inner edges of the sensing electrodes. The A was measured to be 6475mm² while L: 90mm.

Hence:

$$\rho = 0.072R$$